

To whom it concerns,

FertiPro N.V. updated the instructions for use of Fructose Test. A summary of the changes is listed in the table below.


Instructions for use	Fructose Test
Previous version	FP09 I29 R01 C.1
New version	FP09 I29 R01 C.2
Changes	<ul style="list-style-type: none">- Statement on stability after transport removed- Addition of symbol for number of tests- Deletion of '96 tests' in product reference- Reagent 1: change to 2x25ml volume (for more information, see http://fertipro.com/files/Newsitem-IVDR_2023.pdf)- Update in layout of method- All other languages than English: general update in layout

Name: Eline De Reu

Function: Regulatory Affairs Associate

Date: 23/09/2025

Signature and stamp:


Eline De Reu
Regulatory Affairs
FertiPro N.V.