

26/03/2024

To whom it concerns,

FertiPro NV updated the instructions for use of the SpermMar Test IgG. A summary of the changes is listed in the table below.

Instructions for use	SpermMar Test IgG (Positive and Negative Control)
Previous version	FP09 I02 R01 F.1
New version	FP09 I02 R01 F.2
Changes	<ul style="list-style-type: none">- New symbol in the symbol glossary: indication of the number of test- Correction of error: 50 tests can be performed with the SpermMar Test IgG Positive and Negative Control instead of 25 tests- Reference to stability after transport was removed in the IFU. Note however, that stability upon short-term storage at elevated temperatures during transport is still ensured as indicated in the transport stability statement available on our website: https://fertipro.com/quality/

Name: Laura Verspeelt

Function: Regulatory Affairs Associate

Date: 26/03/2024

Signature and stamp:



Laura Verspeelt
Regulatory Affairs
FertiPro N.V.

COMPANY ADDRESS

FertiPro nv
Industriepark Noord 32
B-8730 Beernem / Belgium

tel. +32 (0)50 79 18 05
fax +32 (0)50 79 17 99
info@fertipro.com
www.fertipro.com

RPR Brugge
VAT BE 0448 767 035
IBAN BE 98 7374 4024 9293
BIC KREDBEBB