


09/04/2025

To whom it concerns,

FertiPro NV updated the instructions for use of the SpermMar Test IgG. A summary of the changes is listed in the table below.

Instructions for use	SpermMar Test IgG (Positive and Negative Control)
Previous version	FP09 I02 R01 F.2
New version	FP09 I02 R01 F.3
Changes	<ul style="list-style-type: none">- Small textual changes- Update of the barcode reference to the demonstration video- Extra information on the positive and negative controls

Name: Laura Verspeelt
Function: Regulatory Affairs Associate
Date: 09/04/2025
Signature and stamp:

 Laura Verspeelt
Regulatory Affairs
FertiPro N.V.

COMPANY ADDRESS

FertiPro nv
Industriepark Noord 32
B-8730 Beernem / Belgium

tel. +32 (0)50 79 18 05
fax +32 (0)50 79 17 99
info@fertipro.com
www.fertipro.com

RPR Brugge
VAT BE 0448 767 035
IBAN BE 98 7374 4024 9293
BIC KREDBEBB